

F2 - Acquisition Parameters  
Date\_ 20081202  
Time 19.38  
INSTRUM spect  
PROBHD 5 mm QNP 1H/13  
PULPROG zg30  
TD 65536  
SOLVENT CDCl3  
NS 16  
DS 2  
SWH 8278.146 Hz  
FIDRES 0.126314 Hz  
AQ 3.9584243 sec  
RG 57  
DW 60.400 usec  
DE 6.00 usec  
TE 291.2 K  
D1 1.00000000 sec  
TD0 1

===== CHANNEL f1 =====  
NUC1 1H  
P1 14.00 usec  
PL1 0.00 dB  
SFO1 400.1324710 MHz

F2 - Processing parameters  
SI 65536  
SF 400.1300054 MHz  
WDW EM  
SSB 0  
LB 0.30 Hz  
GB 0  
PC 1.00

