

STANDARD PROTON PARAMETERS

Pulse Sequence: s2pul

Solvent: CDCl3  
Ambient temperature  
File: TPW-II-116A502  
INOVA-500 "zippy"

Relax. delay 0.100 sec  
Pulse 73.1 degrees  
Acq. time 4.999 sec  
Width 12012.0 Hz  
Single scan

OBSERVE H1, 500.4294976 MHz  
DATA PROCESSING  
FT size 262144  
Total time 0 min, 5 sec

